

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 4/9/2015

DBH

fid

Instrument Serial Number: 010484

As of 09-Apr-15

Start Date	End Date	Agency Name
09-Apr-15		New Kent County SO
17-Mar-15	09-Apr-15	DFS Central Lab
07-Oct-14	17-Mar-15	Fluvanna County Sheriff's Office
24-Sep-14	07-Oct-14	DFS Central Lab
23-Apr-14	24-Sep-14	Prince William County PD
08-Apr-14	23-Apr-14	DFS Central Lab
13-Jun-12	08-Apr-14	Grayson County Sheriff's Office
16-Jun-08	13-Jun-12	DFS Central Lab

DBH
ml

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010484 Worksheet Start Date 4/9/2015
Location Henrico County Regional Jail East
Address 17320 New Kent Hwy Barhamsville VA 23011
DFS Technician Donald Hall License No. 21250

☐ Laboratory

☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 763 Reference Barometer (mm HG) 764
Reference Barometer(RB)Serial # 009109 RB Calibration Due 7/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.306
0.300	0.301	0.291	0.310	Sample 2	0.305
Precision		sample min	sample max	Sample 3	0.305
0.001		0.305	0.306		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.101
0.100	0.100	0.097	0.103	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.101
0.001		0.100	0.101		

Dry gas standard Lot No. (with tank no.) AG501301-44

☐ Replaced dry gas standard (+O-ring)

☒ Installed at Location

☐ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:

50

Notes:

Instrument Serial Number

010484

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

4/9/15

Issuing Analyst

Jed
RLN

Date

4/13/15

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010484 Test Number: 811

Test Date: 04/09/2015 Test Time: 14:12 EDT

Dry Gas Target: 0.301

Lot Number: AG428003-05 Exp Date: 10/07/2016

Tank Pressure: 286 psi Barometric Pressure: 763 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	14:12
CHK	0.306	14:13
BLK	0.000	14:15
CHK	0.305	14:16
BLK	0.000	14:18
CHK	0.305	14:18

Test Status: Success

Calibration CRC: 2E06FC2F

Intox EC/IR-II: Accuracy Check

DBH

M

Department of Forensic Science

Serial Number: 010484 Test Number: 812

Test Date: 04/09/2015 Test Time: 14:21 EDT

Dry Gas Target: 0.100

Lot Number: AG501301-44 Exp Date: 01/13/2017

Tank Pressure: 1120 psi Barometric Pressure: 763 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	14:21
CHK	0.101	14:22
BLK	0.000	14:24
CHK	0.100	14:24
BLK	0.000	14:26
CHK	0.101	14:27

Test Status: Success

Calibration CRC: 2E06FC2F

DBH
fid



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--------------------------------------	----------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, P		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 04/09/2015
TEST EQUIPMENT NUMBER 010434		

RESULTS: TIME SAMPLE TAKEN 14:36 EDT
0.00
SAMPLE'S ALCOHOL CONTENT _____ GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 09 Apr 2015 at 16:45

DBH
nd

Test Results

Instrument Serial Number 010484

Test # 000813 Subject Test

Test Location 1 Department of
Test Date 09 Apr 2015

Test Location 2 Forensic Science
Test Time 14:29
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 09 Apr 2015 End Time 14:37

Result Time 14:36

Result Date 09 Apr 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 14:29

Data Type BLK

Sample Value 0.000

Sample Time 14:30

Data Type CHK

Sample Value 0.100

Sample Time 14:31

Data Type BLK

Sample Value 0.000

Sample Time 14:32

Data Type SUBJ

Sample Value 0.000

Sample Time 14:33

Data Type BLK

Sample Value 0.000

Sample Time 14:34

Data Type SUBJ

Sample Value 0.000

Sample Time 14:36

Data Type BLK

Sample Value 0.000

Sample Time 14:37

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG501301-44

Standard Expiration Date 01/13/2017

Tank Pressure 1120

Barometric Pressure 763 mmHg

Blow Sample Number 1 Blow Duration 3.60 sec

Blow Volume 2199 cc End-of-Blow Time 14:33

Blow Sample Number 2 Blow Duration 3.57 sec

Blow Volume 2054 cc End-of-Blow Time 14:36

Tamper Evident Stamp 10296909

Test Status Code 0

Test Status Success